



CLIFT INNOVATIONS PTY. LTD.

ABN: 24 286 188 328
Unit 10/ 13 Walkers Road
Nunawading, Vic. 3131
Phone: 9877-3109
Fax: 9999-1406
Mob: 0411-789702
Web: www.clift1.com
Email: bruce@clift1.com

December 2017

ADCM-HW NURSE CALL SYSTEM



FEATURES:

- | | |
|--|--|
| <ul style="list-style-type: none">• PRE-PROGRAMMED:• EXPANDABLE INPUTS• RESPONSE TIME:• BATTERY BACKED:• LED DISPLAY:• SMS GATEWAY:• PAGING GATEWAY:• AUXILIARY OUTPUTS:• WHITE BOARD DISPLAY:• USB DATA BACKUP:• REPORTING:• ZONE/WARD SWITCHING:• 24V POWER SUPPLY: | <ul style="list-style-type: none">Totally programmed by Clift to Site RequirementsIn modules of 16 latched inputs (Max 1600)Less than 1 secondInternal 12V BatteryAddressable Bestlite single, or double sidedGSM modem (Optional)Local area pager (Optional)16 general purpose outputs, negative switchingHDMI Current call list display & elapsed time8 Gig data storageLocal and CLOUD based (Optional)3 times daily24V 2A Power Supply |
|--|--|